


PDS No. 6531xx	PRODUCT DATA SHEET				Page 1 of 1	
Revision 06	60 Well Microtest Plate (Terasaki Plate), PS					
	Greiner Item-No. 6531xx					
Valid for Item-No.:	653180	653190				

1.	Description / Specification	
1.1	Description	PS Microtest Plate (Terasaki Plate), 60 Well, conical well profile, flat well bottom, alphanumeric well coding, with lid HLA / Microtest Plate, physical surface treatment
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	10 µl
1.4	Material / Resin	Plate and lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Plate and lid: clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for cell culture to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	653180, -190: 4 months
2.7	Other Information	-

3.	Packaging	653180	653190
3.1	Pieces / Bag	10 / bag	120 / folding carton
3.2	Pieces / Box	270	480
3.3	Lot-No.	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	
3.4	Other Information	-	

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 05	Date 26 March 2025	Date 27 March 2025	Date 27 March 2025	
Date 12.08.2024	Name S. Kaelberer	Name Dr. L. Breth	Name A. Mackowski	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.