


PDS No. 6331xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 12	Petri Dish with Vents			
	Greiner Item-No. 6331xx			
Valid for Item-No.:	633102	633161 (sterile)	633180	

1.	Description / Specification	
1.1	Description	Petri dish with vents 633102 (94 x 16 mm): heavy design 633161 (94 x 16 mm): heavy design, sterile 633180 (94 x 16 mm): standard
1.2	Dimensions	See Customer Drawing 633102, -161: total weight: 17.0 – 18.5 g 633180: total weight: 13.0 – 14.7 g
1.3	Volume	Max. volume: 80 ml
1.4	Material / Resin	Dish: PS (Polystyrene), free of heavy metal Lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Dish: clear Lid: clear
1.6	Sterilisation	633102, -180: no 633161: SAL 10 ⁻³
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for bacteriology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	- For single use only - Elevations for ventilation of culture

2.	Features	
2.1	Basic features	Hydrophobic
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	633102, -180: n/a 633161: 5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	20
3.2	Pieces / Box	480
3.3	Lot-No.	633102, -161: F YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.) 633180: G YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 11	Date 5 September 2022	Date 14 September 2022	Date 15 September 2022	
Date 03.05.2019	Name S. Kaelberer	Name Dr. R. Daum	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.