PDS No. 63218x	PRODUCT DATA SHEET	Page 1 of 1
Revision 11	Petri Dish, without Vents, 94 x 16 mm	6
	Greiner Item-No. 6321xx	greiner
Valid for Item-No.:	632180 632181 (sterile)	

1.	Description / Specification			
1.1	Description	Petri Dish without vents, 94 x 16 mm		
		632180: standard design		
		632181: standard design, sterile		
1.2	Dimensions	See Customer Drawing		
		Total weight: 13.0 – 14.7 g		
1.3	Volume	Max. volume: 80 ml		
Working volume: 10		Working volume: 10 – 40 ml		
1.4	Material / Resin	Dish and lid: PS (Polystyrene), free of heavy metal		
1.5	Colour	Dish and lid: clear		
1.6	Sterilisation	632 <i>180</i> : no		
		632181: SAL 10 ⁻³		
1.7	Quality Control	- Raw Material-Control: physical testing		
- Product-Control: testing of attribu		- Product-Control: testing of attributive and variable characteristics in		
		accordance with the valid specification		
1.8	Intended Use	General laboratory product for bacteriology to be used by qualified		
		personnel in a laboratory environment.		
1.9	Other Information	For single use only		

2.	Features	
2.1	Basic features	Hydrophobic
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	632180: n/a 632181: 5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	20
3.2	Pieces / Box	480
3.3	Lot-No.	F YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Data direct cappet to change thin out induce.						
Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this		
Revision	Date	Date	Date	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This		
10	23 January 2025	24 January 2025	24 January 2025	document may not be reproduced for any		
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,		
15.09.2022	S. Kaelberer	T. Binder	Dr. CK. Chai	and copyright are reserved.		