


PDS No. 76707x	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 05	96 Well ELISA-Plate, PS, 12 x U8 Strips, U-Bottom			
	Greiner Item-No. 76707x			
Valid for Item-No.:	<b>767070</b>	<b>767071</b>		

1.	Description / Specification	
1.1	Description	PS Strip plate ELISA, 12 x U8 strips mounted in frame, solid U-bottom (round), alphanumeric well coding 767070: MICROLON 200, medium binding 767071: MICROLON 600, high binding
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	Total volume: 312 µl Working volume: 50 - 280 µl
1.4	Material / Resin	Strips and Frame: PS (Polystyrene), free of heavy metal
1.5	Colour	Frame: white Strips: clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for immunology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	4 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	100
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 21 December 2021	Date 17 January 2022	Date 17 January 2022	
Date 03.12.2014	Name S. Kaelberer	Name R. Daum	Name A. Illig	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.