


PDS No. 1150xx/ 115101	PRODUCT DATA SHEET			Page 1 of 1
Revision 09	Tube, 5 ml, PS			
	Item-No. 1150xx / 115101			
Valid for Item-No.:	115001	115061	115101	

1.	Description / Specification	
1.1	Description	PS tube <i>115001</i> : Immuno Tube, MICROLON®200, medium binding, round bottom <i>115061</i> : Immuno Tube, MICROLON®600, high binding, round bottom <i>115101</i> : without treatment, round bottom Appropriate grip stopper, Cat.-No. 303321
1.2	Dimensions	See Customer Drawing
1.3	Volume	Working volume: 4 ml
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	No
1.7	Quality Control	<u>Raw Material-Control</u> : <i>115101</i> : physical testing <i>115001, -061</i> : physical and immunological testing <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for immunology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	5.800 x g: swinging-bucket rotor 5.800 x g: fixed-angle rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	<i>115001, -061</i> : 4 years <i>115101</i> : n/a
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	250
3.2	Pieces / Box	2.000
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL : Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 08	Date 4 March 2026	Date 5 March 2026	Date 5 March 2026	
Date 25.10.2021	Name S. Kaelberer	Name Dr. L. Breth	Name A. Mackowski	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.